

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075650.D
 Acq On : 08 Dec 2022 16:08
 Operator : JC\MD
 Sample : N5924-03DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INF1222DL

Quant Time: Dec 09 02:09:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	170187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	242700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	46276	50.656	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.320%
35) Dibromofluoromethane	8.171	113	59435	53.824	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.640%
50) Toluene-d8	10.571	98	118761	47.752	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.500%
62) 4-Bromofluorobenzene	12.853	95	88859	48.406	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.820%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.606	83	11688	3.878	ug/l	90
40) Benzene	8.612	78	354548	48.088	ug/l	96
52) Toluene	10.635	92	83249	17.522	ug/l	98
67) Ethyl Benzene	11.965	91	384525	44.968	ug/l	96
68) m/p-Xylenes	12.071	106	455725	131.239	ug/l	99
69) o-Xylene	12.400	106	48983	14.197	ug/l	95
73) Isopropylbenzene	12.700	105	40722	4.594	ug/l	97
78) n-propylbenzene	13.041	91	87444	8.840	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	79175	10.752	ug/l	100
84) 1,2,4-Trimethylbenzene	13.488	105	305370	41.342	ug/l	98
89) n-Butylbenzene	14.059	91	3784	0.585	ug/l #	80
95) Naphthalene	15.641	128	101699	15.609	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075650.D
Acq On : 08 Dec 2022 16:08
Operator : JC\MD
Sample : N5924-03DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INF1222DL

Quant Time: Dec 09 02:09:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

