

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075651.D
 Acq On : 08 Dec 2022 16:32
 Operator : JC\MD
 Sample : N5879-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 09 02:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	273877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	234920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	44031	48.752	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	8.177	113	58509	54.383	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.760%
50) Toluene-d8	10.571	98	116908	48.246	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.500%
62) 4-Bromofluorobenzene	12.853	95	85488	47.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.600%
Target Compounds						
					Qvalue	
40) Benzene	8.612	78	2772	0.386	ug/l	98
52) Toluene	10.630	92	1101	0.238	ug/l	98
67) Ethyl Benzene	11.971	91	5134	0.620	ug/l	90
68) m/p-Xylenes	12.071	106	5922	1.762	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075651.D
Acq On : 08 Dec 2022 16:32
Operator : JC\MD
Sample : N5879-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 09 02:09:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

