

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075652.D
 Acq On : 08 Dec 2022 16:56
 Operator : JC\MD
 Sample : N5879-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 09 02:09:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	164035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262181	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	226844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	46733	53.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.140%	
35) Dibromofluoromethane	8.177	113	60592	58.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.660%	
50) Toluene-d8	10.571	98	121171	52.236	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.480%	
62) 4-Bromofluorobenzene	12.853	95	87677	51.208	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075652.D
Acq On : 08 Dec 2022 16:56
Operator : JC\MD
Sample : N5879-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 09 02:09:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

