

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075657.D
 Acq On : 08 Dec 2022 19:20
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 09 02:30:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	221381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94812	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	43468	51.800	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.600%
35) Dibromofluoromethane	8.171	113	56052	56.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.260%
50) Toluene-d8	10.571	98	109614	48.735	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.480%
62) 4-Bromofluorobenzene	12.853	95	75763	45.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075657.D
Acq On : 08 Dec 2022 19:20
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 09 02:30:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

