

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
 Data File : VN080403.D
 Acq On : 08 Dec 2023 13:49
 Operator : JC\MD
 Sample : VN1208MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1208MBL01

Quant Time: Dec 08 22:37:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	145657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	287670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125775	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132252	54.990	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.980%
35) Dibromofluoromethane	8.171	113	103990	53.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.140%
50) Toluene-d8	10.571	98	349196	49.900	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.853	95	135198	51.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
Data File : VN080403.D
Acq On : 08 Dec 2023 13:49
Operator : JC\MD
Sample : VN1208MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1208MBL01

Quant Time: Dec 08 22:37:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

