

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
 Data File : VN080416.D
 Acq On : 08 Dec 2023 19:08
 Operator : JC\MD
 Sample : 05718-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-LIQUID-20231206

Quant Time: Dec 08 22:45:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	267653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	498463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	514735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	247828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	201548	45.606	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.220%		
35) Dibromofluoromethane	8.165	113	177201	48.179	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.360%		
50) Toluene-d8	10.571	98	658057	49.633	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.260%		
62) 4-Bromofluorobenzene	12.853	95	271999	54.314	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	108.620%		
Target Compounds						
16) Acetone	4.442	43	13714	9.239	ug/l	94
25) 2-Butanone	7.494	43	5259	2.982	ug/l	88
30) Chloroform	7.971	83	68044	10.615	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
Data File : VN080416.D
Acq On : 08 Dec 2023 19:08
Operator : JC\MD
Sample : 05718-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-LIQUID-20231206

Quant Time: Dec 08 22:45:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

