

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
 Data File : VN080417.D
 Acq On : 08 Dec 2023 19:32
 Operator : JC\MD
 Sample : PB157593TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157593TB

Quant Time: Dec 08 22:45:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	317569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	578428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	594344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	239435	45.663	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.320%		
35) Dibromofluoromethane	8.171	113	208885	48.942	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.880%		
50) Toluene-d8	10.571	98	771092	50.118	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.240%		
62) 4-Bromofluorobenzene	12.853	95	301638	51.905	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.820%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.518	59	9805	17.004	ug/l	97
16) Acetone	4.430	43	38603	21.919	ug/l	98
20) Methylene Chloride	5.271	84	24590	6.667	ug/l #	89
25) 2-Butanone	7.488	43	48833	23.337	ug/l	88
37) Ethyl Acetate	7.571	43	6235	1.455	ug/l #	95
43) Isopropyl Acetate	8.688	43	207506	24.878	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
Data File : VN080417.D
Acq On : 08 Dec 2023 19:32
Operator : JC\MD
Sample : PB157593TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157593TB

Quant Time: Dec 08 22:45:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

