

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
 Data File : VN080421.D
 Acq On : 08 Dec 2023 21:07
 Operator : JC\MD
 Sample : 05671-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-35

Quant Time: Dec 08 22:47:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	289977	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	542794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	585864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	257787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	219093	45.759	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.520%
35) Dibromofluoromethane	8.165	113	188460	47.055	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	10.571	98	746212	51.685	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.380%
62) 4-Bromofluorobenzene	12.853	95	313525	57.492	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.980%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.518	59	9141	17.361	ug/l #	94
16) Acetone	4.436	43	34476	21.438	ug/l	98
20) Methylene Chloride	5.283	84	31646	9.397	ug/l	95
25) 2-Butanone	7.488	43	38274	20.031	ug/l #	81
43) Isopropyl Acetate	8.688	43	156986	20.057	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120823\
Data File : VN080421.D
Acq On : 08 Dec 2023 21:07
Operator : JC\MD
Sample : 05671-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-35

Quant Time: Dec 08 22:47:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

