

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065015.D
 Acq On : 9 Dec 2020 13:01
 Operator : JC/MD
 Sample : PB133472TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472TB

Quant Time: Dec 10 04:34:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	167828	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	312769	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143365	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.98	65	108646	55.68	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	111.36%		
35) Dibromofluoromethane	7.55	113	100132	53.50	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.00%		
50) Toluene-d8	10.06	98	410397	53.67	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.34%		
62) 4-Bromofluorobenzene	12.38	95	152581	55.75	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.50%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	6261	9.762 ug/l	98
20) Methylene Chloride	4.50	84	5000	2.488 ug/l	96
43) Isopropyl Acetate	8.13	43	15986	3.558 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065015.D
Acq On : 9 Dec 2020 13:01
Operator : JC/MD
Sample : PB133472TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472TB

Quant Time: Dec 10 04:34:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

