

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065016.D
 Acq On : 9 Dec 2020 13:25
 Operator : JC/MD
 Sample : PB133473ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133473ZHE#17

Manual Integrations
 APPROVED

MMDadoda
 12/10/2020 9:14:35 AM

Quant Time: Dec 10 04:35:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	166374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	308514	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	317571	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136103	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105997	54.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	109.60%		
35) Dibromofluoromethane	7.55	113	99327	53.81	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.62%		
50) Toluene-d8	10.06	98	381934	50.64	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.28%		
62) 4-Bromofluorobenzene	12.38	95	140891	52.19	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.38%		
Target Compounds						
16) Acetone	3.78	43	7699	12.109	ug/l	97
20) Methylene Chloride	4.51	84	2219m	1.114	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065016.D
Acq On : 9 Dec 2020 13:25
Operator : JC/MD
Sample : PB133473ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133473ZHE#17

Manual Integrations
APPROVED

MMDadoda
12/10/2020 9:14:35 AM

Quant Time: Dec 10 04:35:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

