

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065017.D
 Acq On : 9 Dec 2020 13:49
 Operator : JC/MD
 Sample : PB133473ZHE#18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133473ZHE#18

Quant Time: Dec 10 04:36:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	161776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	299010	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	306341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	133699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	103278	54.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	109.82%		
35) Dibromofluoromethane	7.55	113	98700	55.16	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	110.32%		
50) Toluene-d8	10.06	98	374057	51.17	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.34%		
62) 4-Bromofluorobenzene	12.38	95	140376	53.66	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	107.32%		
Target Compounds						
16) Acetone	3.78	43	7741	12.521	ug/l	91
20) Methylene Chloride	4.50	84	2643	1.364	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065017.D
Acq On : 9 Dec 2020 13:49
Operator : JC/MD
Sample : PB133473ZHE#18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133473ZHE#18

Quant Time: Dec 10 04:36:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

