

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065020.D
 Acq On : 9 Dec 2020 15:02
 Operator : JC/MD
 Sample : PB133472ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#01

Quant Time: Dec 10 04:39:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	300438	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	331722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136992	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	105541	56.55	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	113.10%	
35) Dibromofluoromethane	7.55	113	99159	55.16	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	110.32%	
50) Toluene-d8	10.06	98	394274	53.68	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	107.36%	
62) 4-Bromofluorobenzene	12.38	95	146322	55.66	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	111.32%	

Target Compounds

					Qvalue
16) Acetone	3.78	43	5344	8.711 ug/l	93
20) Methylene Chloride	4.51	84	5255	2.734 ug/l	91
43) Isopropyl Acetate	8.13	43	14944	3.462 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065020.D
Acq On : 9 Dec 2020 15:02
Operator : JC/MD
Sample : PB133472ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#01

Quant Time: Dec 10 04:39:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

