

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065021.D
 Acq On : 9 Dec 2020 15:26
 Operator : JC/MD
 Sample : PB133472ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#02

Quant Time: Dec 10 04:41:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	165040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	304821	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	336780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136609	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	108320	56.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.90%	
35) Dibromofluoromethane		7.55	113	98251	53.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.74%	
50) Toluene-d8		10.06	98	400160	53.69	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.38%	
62) 4-Bromofluorobenzene		12.38	95	148801	55.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5752	9.120	ug/l	100
20) Methylene Chloride	4.50	84	5160	2.611	ug/l	91
43) Isopropyl Acetate	8.13	43	14986	3.422	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065021.D
Acq On : 9 Dec 2020 15:26
Operator : JC/MD
Sample : PB133472ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#02

Quant Time: Dec 10 04:41:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

