

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065022.D
 Acq On : 9 Dec 2020 15:50
 Operator : JC/MD
 Sample : PB133472ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#03

Quant Time: Dec 10 04:42:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	312561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	331228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136758	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	106649	55.63	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.26%
35) Dibromofluoromethane	7.55	113	99121	53.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.00%
50) Toluene-d8	10.06	98	402315	52.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.30%
62) 4-Bromofluorobenzene	12.38	95	147940	54.09	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.18%

Target Compounds

					Qvalue
16) Acetone	3.78	43	5918	9.391 ug/l	98
20) Methylene Chloride	4.51	84	5247	2.657 ug/l	89
43) Isopropyl Acetate	8.13	43	16087	3.582 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065022.D
Acq On : 9 Dec 2020 15:50
Operator : JC/MD
Sample : PB133472ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#03

Quant Time: Dec 10 04:42:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

