

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065024.D
 Acq On : 9 Dec 2020 16:38
 Operator : JC/MD
 Sample : PB133472ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#05

Quant Time: Dec 10 04:45:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	318760	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	334824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140325	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	107764	56.38	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	112.76%	
35) Dibromofluoromethane	7.55	113	98953	51.88	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	103.76%	
50) Toluene-d8	10.06	98	403028	51.71	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	103.42%	
62) 4-Bromofluorobenzene	12.38	95	149660	53.66	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	107.32%	

Target Compounds

					Qvalue	
16) Acetone	3.78	43	5772	9.187	ug/l	94
20) Methylene Chloride	4.51	84	5361	2.723	ug/l #	91
43) Isopropyl Acetate	8.13	43	18893	4.125	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065024.D
Acq On : 9 Dec 2020 16:38
Operator : JC/MD
Sample : PB133472ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#05

Quant Time: Dec 10 04:45:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

