

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065027.D
 Acq On : 9 Dec 2020 17:50
 Operator : JC/MD
 Sample : PB133472ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#08

Quant Time: Dec 10 04:50:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	158296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311100	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	327258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136011	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	104533	56.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	113.60%		
35) Dibromofluoromethane	7.55	113	97394	52.32	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.64%		
50) Toluene-d8	10.06	98	387752	50.98	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.96%		
62) 4-Bromofluorobenzene	12.38	95	143699	52.79	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	105.58%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	6115	10.108 ug/l	96
20) Methylene Chloride	4.50	84	5166	2.725 ug/l	93
43) Isopropyl Acetate	8.13	43	17703	3.961 ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065027.D
Acq On : 9 Dec 2020 17:50
Operator : JC/MD
Sample : PB133472ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#08

Quant Time: Dec 10 04:50:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

