

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065029.D
 Acq On : 9 Dec 2020 18:38
 Operator : JC/MD
 Sample : PB133472ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#10

Quant Time: Dec 10 04:53:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	161585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	316156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	328899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	137021	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	105315	56.06	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.12%	
35) Dibromofluoromethane		7.55	113	95716	50.60	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.20%	
50) Toluene-d8		10.06	98	391178	50.61	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.22%	
62) 4-Bromofluorobenzene		12.38	95	145478	52.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5938	9.616	ug/l	89
20) Methylene Chloride	4.51	84	5121	2.646	ug/l	94
43) Isopropyl Acetate	8.13	43	19406	4.272	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065029.D
Acq On : 9 Dec 2020 18:38
Operator : JC/MD
Sample : PB133472ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#10

Quant Time: Dec 10 04:53:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

