

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065030.D
 Acq On : 9 Dec 2020 19:02
 Operator : JC/MD
 Sample : PB133472ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#11

Quant Time: Dec 10 04:54:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	162108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	324632	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	337083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136758	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	108410	57.52	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.04%	
35) Dibromofluoromethane		7.55	113	99651	51.30	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.60%	
50) Toluene-d8		10.06	98	404390	50.95	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.90%	
62) 4-Bromofluorobenzene		12.38	95	147062	51.77	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5668	9.149	ug/l	97
20) Methylene Chloride	4.51	84	4808	2.477	ug/l	90
43) Isopropyl Acetate	8.13	43	18352	3.935	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065030.D
Acq On : 9 Dec 2020 19:02
Operator : JC/MD
Sample : PB133472ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#11

Quant Time: Dec 10 04:54:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

