

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065033.D
 Acq On : 9 Dec 2020 20:15
 Operator : JC/MD
 Sample : PB133472ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472ZHE#14

Quant Time: Dec 10 05:00:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	151813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	300502	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	320786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129089	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	100430	56.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.80%
35) Dibromofluoromethane	7.55	113	93981	52.27	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.54%
50) Toluene-d8	10.06	98	379176	51.61	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.22%
62) 4-Bromofluorobenzene	12.38	95	139251	52.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.92%

Target Compounds

					Qvalue
16) Acetone	3.78	43	5145	8.868 ug/l	99
20) Methylene Chloride	4.51	84	5000	2.750 ug/l #	87
43) Isopropyl Acetate	8.13	43	17645	4.087 ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065033.D
Acq On : 9 Dec 2020 20:15
Operator : JC/MD
Sample : PB133472ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472ZHE#14

Quant Time: Dec 10 05:00:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

