

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120920\
 Data File : VN065015.D
 Acq On : 9 Dec 2020 13:01
 Operator : JC/MD
 Sample : PB133472TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133472TB

Quant Time: Dec 10 04:34:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	167828	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	312769	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143365	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.98	65	108646	55.68	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	111.36%		
35) Dibromofluoromethane	7.55	113	100132	53.50	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.00%		
50) Toluene-d8	10.06	98	410397	53.67	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.34%		
62) 4-Bromofluorobenzene	12.38	95	152581	55.75	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.50%		

Target Compounds


					Qvalue
16) Acetone	3.78	43	6261	9.762 ug/l	98
20) Methylene Chloride	4.50	84	5000	2.488 ug/l	96
43) Isopropyl Acetate	8.13	43	15986	3.558 ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120920\
Data File : VN065015.D
Acq On : 9 Dec 2020 13:01
Operator : JC/MD
Sample : PB133472TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133472TB

Quant Time: Dec 10 04:34:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN065015.D

Acq: 9 Dec 2020 13:01

Instrument :

MSVOA_N

ClientSampleId :


PB133472TB

Tot Ion:168 Resp: 167828

Ion Ratio Lower Upper

168 100

99 53.3 43.0 64.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.63

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#16

Acetone

Concen: 9.762 ug/l

RT: 3.78 min Scan# 636

Delta R.T. 0.01 min

Lab File: VN065015.D


Acq: 9 Dec 2020 13:01

Tgt Ion: 43 Resp: 6261

Ion Ratio Lower Upper

43 100

58 31.2 25.7 38.5



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

3.78

2000


1500


1000

500

0

Time--> 3.70 3.75 3.80 3.85









#20
Methylene Chloride
Concen: 2.488 ug/l
RT: 4.50 min Scan# 859
Delta R.T. 0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01

Instrument :
MSVOA_N
ClientSampled :
PB133472TB

Tot Ion: 84 Resp: 5000
Ion Ratio Lower Upper
84 100
49 115.4 93.4 140.0
51 35.5 28.2 42.4
86 74.3 51.7 77.5




Abundance Ion 83.90 (83.60 to 84.60): VNC
Ion 48.90 (48.60 to 49.60): VNC
Ion 50.90 (50.60 to 51.60): VNC
Ion 85.90 (85.60 to 86.60): VNC






#33
1,2-Dichloroethane-d4
Concen: 55.684 ug/l
RT: 7.98 min Scan# 1942
Delta R.T. -0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01

Tgt Ion: 65 Resp: 108646
Ion Ratio Lower Upper
65 100
67 53.4 0.0 108.2



Abundance Ion 64.90 (64.60 to 65.60): VNC
Ion 66.90 (66.60 to 67.60): VNC









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.55 min Scan# 2118
Delta R.T. 0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01

Instrument :
MSVOA_N
ClientSampleId :
PB133472TB

Tot Ion:114 Resp: 312769
Ion Ratio Lower Upper
114 100
63 19.4 0.0 38.6
88 15.6 0.0 32.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): V
Ion 87.90 (87.60 to 88.60): V






#35
Dibromofluoromethane
Concen: 53.503 ug/l
RT: 7.55 min Scan# 1807
Delta R.T. 0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01

Tgt Ion:113 Resp: 100132
Ion Ratio Lower Upper
113 100
111 101.7 81.6 122.4
192 20.3 16.2 24.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.558 ug/l

RT: 8.13 min Scan# 1986

Delta R.T. -0.00 min

Lab File: VN065015.D

Acq: 9 Dec 2020 13:01

Instrument :

MSVOA_N

ClientSampleId :

PB133472TB


Tot Ion: 43 Resp: 15986

Ion Ratio Lower Upper

43 100

61 19.3 17.2 25.8




87 11.5 11.4 17.0



Abundance Ion 43.00 (42.70 to 43.70): VN064779.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN064779.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN064779.D (-1968) (-)



#50

Toluene-d8

Concen: 53.669 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN065015.D


Acq: 9 Dec 2020 13:01

Tgt Ion: 98 Resp: 410397

Ion Ratio Lower Upper



98 100


100 64.1 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN064779.D (-2572) (-)

Ion 100.00 (99.70 to 100.70): VN064779.D (-2572) (-)









#62
4-Bromofluorobenzene
Concen: 55.755 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01




Instrument :
MSVOA_N
ClientSampleId :
PB133472TB


Tot Ion: 95 Resp: 152581
Ion Ratio Lower Upper
95 100
174 83.2 0.0 160.4
176 79.3 0.0 158.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN065015.D
Acq: 9 Dec 2020 13:01

Tgt Ion: 117 Resp: 341159
Ion Ratio Lower Upper
117 100
82 54.8 44.2 66.2
119 32.3 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3600

Delta R.T. -0.00 min

Lab File: VN065015.D

Acq: 9 Dec 2020 13:01

Instrument :

MSVOA_N

ClientSampleId :

PB133472TB


Tot Ion:152 Resp: 143365

Ion Ratio Lower Upper

152 100

115 57.5 29.7 89.1

150 156.7 0.0 350.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

