


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN120920\
 Data File : VN065009.D
 Acq On : 9 Dec 2020 9:21
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Title : SW846 8260
 Last Update : Mon Nov 23 13:54:12 2020



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	4799	PASS
75	95	30	60	47.7	13114	PASS
95	95	100	100	100.0	27469	PASS
96	95	5	9	7.1	1952	PASS
173	174	0.00	2	1.5	349	PASS
174	95	50	100	84.7	23269	PASS
175	174	5	9	7.9	1846	PASS
176	174	95	101	95.3	22178	PASS
177	176	5	9	7.4	1650	PASS