

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069940.D
 Acq On : 10 Dec 2021 4:12
 Operator : JC/MD
 Sample : PB141262ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141262ZHE#06

Quant Time: Dec 10 08:14:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.088	168	1136020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2192248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2516609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1014148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	912160	52.004	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.000%			
35) Dibromofluoromethane	8.024	113	675787	49.213	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.420%			
50) Toluene-d8	10.443	98	2925372	53.056	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.120%			
62) 4-Bromofluorobenzene	12.731	95	1203718	60.261	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 120.520%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.315	59	2475	0.746	ug/l #	77
13) Acrolein	4.052	56	1049	0.372	ug/l #	48
16) Acetone	4.296	43	70766	6.707	ug/l	92
18) Methyl Acetate	4.865	43	47789	1.501	ug/l	92
20) Methylene Chloride	5.101	84	40337	2.876	ug/l #	85
25) 2-Butanone	7.343	43	13333	0.930	ug/l #	73
29) Tetrahydrofuran	7.694	42	4223	0.485	ug/l #	52
31) Cyclohexane	8.083	56	29478	1.208	ug/l #	18
36) 1,1-Dichloropropene	8.088	75	98572	4.723	ug/l #	49
37) Ethyl Acetate	7.343	43	13333	0.429	ug/l #	68
38) Carbon Tetrachloride	8.086	117	120349	5.881	ug/l #	16
43) Isopropyl Acetate	8.566	43	177148	3.854	ug/l #	78
48) Methyl methacrylate	9.547	41	33071	1.542	ug/l #	8
49) 1,4-Dioxane	9.614	88	479	1.244	ug/l #	27
51) 4-Methyl-2-Pentanone	10.336	43	19981	0.670	ug/l	93
53) t-1,3-Dichloropropene	10.727	75	96	2.798	ug/l #	1
54) cis-1,3-Dichloropropene	10.245	75	171535	8.131	ug/l #	72
56) Ethyl methacrylate	10.878	69	1770	2.336	ug/l #	1
57) 1,3-Dichloropropane	11.089	76	104111	3.763	ug/l #	43
58) 2-Chloroethyl Vinyl ether	9.901	63	27	18.392	ug/l #	45
59) 2-Hexanone	11.089	43	1891011	83.831	ug/l #	47
60) Dibromochloromethane	11.492	129	862	2.425	ug/l #	11
74) N-amyl acetate	12.197	43	74253	1.861	ug/l #	52
76) 1,2,3-Trichloropropane	12.731	75	671451	23.087	ug/l #	1
80) 1,3,5-Trimethylbenzene	13.061	105	18753	0.258	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.731	75	671451	73.485	ug/l #	5
93) 1,2,4-Trichlorobenzene	15.268	180	639	2.914	ug/l #	50
95) Naphthalene	15.509	128	4086	4.061	ug/l #	88
96) 1,2,3-Trichlorobenzene	15.705	180	577	2.469	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069940.D
Acq On : 10 Dec 2021 4:12
Operator : JC/MD
Sample : PB141262ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06

Quant Time: Dec 10 08:14:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 21
Delta R.T. 0.002 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion:168 Resp: 1136020
Ion Ratio Lower Upper
168 100
99 61.4 41.3 61.9



#11
Tert butyl alcohol
Concen: 0.746 ug/l
RT: 5.315 min Scan# 1240
Delta R.T. -0.075 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 59 Resp: 2475
Ion Ratio Lower Upper
59 100
57 2.3 8.7 13.1#









#13
Acrolein
Concen: 0.372 ug/l
RT: 4.052 min Scan# 769
Delta R.T. 0.014 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 56 Resp: 1049
Ion Ratio Lower Upper
56 100
55 27.6 56.6 85.0#



#16
Acetone
Concen: 6.707 ug/l
RT: 4.296 min Scan# 860
Delta R.T. 0.005 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 43 Resp: 70766
Ion Ratio Lower Upper
43 100
58 24.9 23.1 34.7









#18
Methyl Acetate
Concen: 1.501 ug/l
RT: 4.865 min Scan# 1160
Delta R.T. 0.008 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 43 Resp: 47789
Ion Ratio Lower Upper
43 100
74 20.1 19.4 29.2



#20
Methylene Chloride
Concen: 2.876 ug/l
RT: 5.101 min Scan# 1160
Delta R.T. 0.008 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 84 Resp: 40337
Ion Ratio Lower Upper
84 100
49 143.7 93.5 140.3#
51 43.4 29.4 44.2
86 64.7 53.0 79.4









#25
2-Butanone
Concen: 0.930 ug/l
RT: 7.343 min Scan# 1994
Delta R.T. 0.005 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 43 Resp: 13333
Ion Ratio Lower Upper
43 100
72 13.4 21.7 32.5#



#29
Tetrahydrofuran
Concen: 0.485 ug/l
RT: 7.694 min Scan# 2127
Delta R.T. 0.008 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12


Tgt Ion: 42 Resp: 4223
Ion Ratio Lower Upper
42 100
72 6.1 36.3 54.5#
71 20.7 34.7 52.1#








#31
Cyclohexane
Concen: 1.208 ug/l
RT: 8.083 min Scan# 21
Delta R.T. -0.011 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06







Tgt Ion: 56 Resp: 29478
Ion Ratio Lower Upper
56 100
69 150.4 27.1 40.7#
84 127.4 73.4 110.0#



#33
1,2-Dichloroethane-d4
Concen: 52.004 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. 0.003 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 65 Resp: 912160
Ion Ratio Lower Upper
65 100
67 51.0 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.971 min Scan# 2

Delta R.T. 0.003 min

Lab File: VN069940.D

Acq: 10 Dec 2021 4:12

Instrument :

MSVOA_N

ClientSampleId :

PB141262ZHE#06





Tgt Ion:114 Resp: 2192248

Ion Ratio Lower Upper

114 100

63 23.6 0.0 37.8

88 15.9 0.0 27.4



#35

Dibromofluoromethane

Concen: 49.213 ug/l

RT: 8.024 min Scan# 2250

Delta R.T. 0.000 min

Lab File: VN069940.D

Acq: 10 Dec 2021 4:12




Tgt Ion:113 Resp: 675787


Ion Ratio Lower Upper

113 100

111 103.0 82.2 123.2

192 18.7 16.9 25.3









#36
1,1-Dichloropropene
Concen: 4.723 ug/l
RT: 8.088 min Scan# 21
Delta R.T. -0.134 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 75 Resp: 98572
Ion Ratio Lower Upper
75 100
110 8.8 18.4 55.2#
77 0.0 24.4 36.6#



#37
Ethyl Acetate
Concen: 0.429 ug/l
RT: 7.343 min Scan# 1996
Delta R.T. -0.072 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12


Tgt Ion: 43 Resp: 13333
Ion Ratio Lower Upper
43 100
61 0.0 11.5 17.3#
70 0.0 7.9 11.9#








#38
Carbon Tetrachloride
Concen: 5.881 ug/l
RT: 8.086 min Scan# 21
Delta R.T. -0.121 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06







Tgt Ion:117 Resp: 120349
Ion Ratio Lower Upper
117 100
119 4.6 76.1 114.1#
121 0.0 24.2 36.2#



#43
Isopropyl Acetate
Concen: 3.854 ug/l
RT: 8.566 min Scan# 2452
Delta R.T. 0.003 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 43 Resp: 177148
Ion Ratio Lower Upper
43 100
61 12.8 20.4 30.6#
87 7.7 11.2 16.8#









#48
Methyl methacrylate
Concen: 1.542 ug/l
RT: 9.547 min Scan# 2818
Delta R.T. -0.014 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 41 Resp: 33071
Ion Ratio Lower Upper
41 100
69 0.2 72.4 108.6#
39 0.0 58.4 87.6#



#49
1,4-Dioxane
Concen: 1.244 ug/l
RT: 9.614 min Scan# 2843
Delta R.T. 0.043 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 88 Resp: 479
Ion Ratio Lower Upper
88 100
43 0.0 23.4 35.2#
58 0.0 50.6 75.8#









#50
Toluene-d8
Concen: 53.056 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 98 Resp: 2925372
Ion Ratio Lower Upper
98 100
100 63.6 50.4 75.6



#51
4-Methyl-2-Pentanone
Concen: 0.670 ug/l
RT: 10.336 min Scan# 3112
Delta R.T. 0.003 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12


Tgt Ion: 43 Resp: 19981
Ion Ratio Lower Upper
43 100
58 42.3 30.4 45.6








#53
t-1,3-Dichloropropene
Concen: 2.798 ug/l
RT: 10.727 min Scan# 31
Delta R.T. 0.008 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06







Tgt Ion: 75 Resp: 96
Ion Ratio Lower Upper
75 100
77 87.0 25.6 38.4#



#54
cis-1,3-Dichloropropene
Concen: 8.131 ug/l
RT: 10.245 min Scan# 3078
Delta R.T. 0.054 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 75 Resp: 171535
Ion Ratio Lower Upper
75 100
77 0.4 26.4 39.6#
39 46.3 42.5 63.7









#56
Ethyl methacrylate
Concen: 2.336 ug/l
RT: 10.878 min Scan# 31
Delta R.T. 0.116 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 69 Resp: 1770
Ion Ratio Lower Upper
69 100
41 3964.8 51.4 77.0#
39 1937.2 34.6 52.0#



#57
1,3-Dichloropropane
Concen: 3.763 ug/l
RT: 11.089 min Scan# 3393
Delta R.T. 0.043 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 76 Resp: 104111
Ion Ratio Lower Upper
76 100
78 0.5 25.7 38.5#









#58
2-Chloroethyl Vinyl ether
Concen: 18.392 ug/l
RT: 9.901 min Scan# 2950
Delta R.T. -0.142 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 63 Resp: 27
Ion Ratio Lower Upper
63 100
106 0.0 23.9 35.9#



#59
2-Hexanone
Concen: 83.831 ug/l
RT: 11.089 min Scan# 3393
Delta R.T. 0.002 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 43 Resp: 1891011
Ion Ratio Lower Upper
43 100
58 16.2 26.9 80.7#









#60
Dibromochloromethane
Concen: 2.425 ug/l
RT: 11.492 min Scan# 311
Delta R.T. 0.252 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion:129 Resp: 862
Ion Ratio Lower Upper
129 100
127 0.0 38.6 116.0#



#62
4-Bromofluorobenzene
Concen: 60.261 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 95 Resp: 1203718
Ion Ratio Lower Upper
95 100
174 76.0 0.0 187.2
176 73.4 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3637
Delta R.T. -0.002 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion:117 Resp: 2516609
Ion Ratio Lower Upper
117 100
82 59.1 42.4 63.6
119 31.9 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.675 min Scan# 4357
Delta R.T. -0.003 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion:152 Resp: 1014148
Ion Ratio Lower Upper
152 100
115 61.1 27.5 82.5
150 158.0 0.0 344.4









#74
N-amy1 acetate
Concen: 1.861 ug/l
RT: 12.197 min Scan# 31
Delta R.T. -0.190 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion: 43	Resp: 74253
Ion Ratio	Lower Upper
43 100	
70 2.3	33.4 50.0#
55 4.9	19.3 28.9#
61 8.8	18.5 27.7#



#76
1,2,3-Trichloropropane
Concen: 23.087 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.150 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 75	Resp: 671451
Ion Ratio	Lower Upper
75 100	
77 1.1	82.6 247.8#









#80
1,3,5-Trimethylbenzene
Concen: 0.258 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. 0.000 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion:105 Resp: 18753
Ion Ratio Lower Upper
105 100
120 49.0 24.6 73.6



#81
trans-1,4-Dichloro-2-butene
Concen: 73.485 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. 0.105 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion: 75 Resp: 671451
Ion Ratio Lower Upper
75 100
53 0.1 90.5 135.7#
89 0.0 38.2 57.4#









#93
1,2,4-Trichlorobenzene
Concen: 2.914 ug/l
RT: 15.268 min Scan# 4950
Delta R.T. 0.003 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12




Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#06


Tgt Ion:180 Resp: 639
Ion Ratio Lower Upper
180 100
182 49.1 48.3 144.9
145 0.0 14.5 43.5#



#95
Naphthalene
Concen: 4.061 ug/l
RT: 15.509 min Scan# 5041
Delta R.T. 0.002 min
Lab File: VN069940.D
Acq: 10 Dec 2021 4:12

Tgt Ion:128 Resp: 4086
Ion Ratio Lower Upper
128 100
127 12.4 10.5 15.7
129 2.1 8.7 13.1#





#96

1,2,3-Trichlorobenzene

Concen: 2.469 ug/l

RT: 15.705 min Scan# 5114

Delta R.T. 0.003 min

Lab File: VN069940.D


Acq: 10 Dec 2021 4:12

Instrument :

MSVOA_N

ClientSampleId :

PB141262ZHE#06



Tgt Ion:180 Resp: 577

Ion Ratio Lower Upper

180 100

182 88.2 47.4 142.2

145 0.0 15.0 45.1#

