

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069914.D
 Acq On : 9 Dec 2021 16:21
 Operator : JC/MD
 Sample : M4940-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-24

Quant Time: Dec 10 06:17:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1225781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2264210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2537913	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1037491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	967685	51.130	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.260%	
35) Dibromofluoromethane	8.021	113	705232	49.725	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.460%	
50) Toluene-d8	10.441	98	2962709	52.025	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.060%	
62) 4-Bromofluorobenzene	12.731	95	1236009	59.911	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 119.820%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.380	59	253991	70.997	ug/l	100
16) Acetone	4.291	43	59157	5.196	ug/l #	85
19) Methyl tert-butyl Ether	5.618	73	96825	2.133	ug/l #	1
22) Diisopropyl ether	6.498	45	52255	1.016	ug/l	91
68) m/p-Xylenes	11.948	106	9328	0.274	ug/l	100
95) Naphthalene	15.509	128	11491	4.194	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069914.D
Acq On : 9 Dec 2021 16:21
Operator : JC/MD
Sample : M4940-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-24

Quant Time: Dec 10 06:17:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

