

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069926.D
 Acq On : 9 Dec 2021 21:26
 Operator : JC/MD
 Sample : M4970-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18

Quant Time: Dec 10 06:23:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1230555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2291113	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2578306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1048773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	952076	50.110	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.220%	
35) Dibromofluoromethane	8.024	113	697021	48.569	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.140%	
50) Toluene-d8	10.440	98	3026592	52.523	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.040%	
62) 4-Bromofluorobenzene	12.731	95	1233157	59.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 118.140%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.374	59	608468	169.423	ug/l	99
19) Methyl tert-butyl Ether	5.621	73	1604801	35.211	ug/l #	87
22) Diisopropyl ether	6.500	45	236168	4.575	ug/l #	89
39) Methylcyclohexane	9.461	83	13042	0.503	ug/l #	85
40) Benzene	8.461	78	2753860	41.308	ug/l	100
67) Ethyl Benzene	11.843	91	366210	3.854	ug/l	98
68) m/p-Xylenes	11.958	106	223964	6.474	ug/l	93
69) o-Xylene	12.280	106	11992	0.344	ug/l	90
73) Isopropylbenzene	12.578	105	76606	0.821	ug/l	98
78) n-propylbenzene	12.921	91	76011	0.732	ug/l	93
80) 1,3,5-Trimethylbenzene	13.058	105	58089	0.773	ug/l	96
84) 1,2,4-Trimethylbenzene	13.369	105	168079	2.350	ug/l	96
95) Naphthalene	15.507	128	70589	5.253	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069926.D
Acq On : 9 Dec 2021 21:26
Operator : JC/MD
Sample : M4970-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18

Quant Time: Dec 10 06:23:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

