

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069927.D
 Acq On : 9 Dec 2021 21:51
 Operator : JC/MD
 Sample : M4970-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-19

Quant Time: Dec 10 06:24:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1199673	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2216500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2489338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1032110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	923303	49.846	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.700%
35) Dibromofluoromethane	8.024	113	676663	48.738	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.480%
50) Toluene-d8	10.440	98	2910564	52.210	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.731	95	1227049	60.757	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.520%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.374	59	622375	177.756	ug/l	100
16) Acetone	4.293	43	22616	2.030	ug/l	94
19) Methyl tert-butyl Ether	5.618	73	8627999	194.182	ug/l	96
22) Diisopropyl ether	6.498	45	102559	2.038	ug/l	88
95) Naphthalene	15.504	128	30708	4.545	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069927.D
Acq On : 9 Dec 2021 21:51
Operator : JC/MD
Sample : M4970-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-19

Quant Time: Dec 10 06:24:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

