

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069935.D
 Acq On : 10 Dec 2021 2:05
 Operator : JC/MD
 Sample : PB141262ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141262ZHE#01

Quant Time: Dec 10 08:12:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1166885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2194781	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2497412	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1011079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	919065	51.012	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 102.020%		
35) Dibromofluoromethane	8.021	113	670258	48.754	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 97.500%		
50) Toluene-d8	10.443	98	2909186	52.701	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 105.400%		
62) 4-Bromofluorobenzene	12.731	95	1198407	59.926	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 119.860%		
Target Compounds						Qvalue
16) Acetone	4.299	43	86658	7.996	ug/l	95
20) Methylene Chloride	5.098	84	36958	2.565	ug/l	87
25) 2-Butanone	7.345	43	18787	1.276	ug/l	91
43) Isopropyl Acetate	8.566	43	181040	3.934	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069935.D
Acq On : 10 Dec 2021 2:05
Operator : JC/MD
Sample : PB141262ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141262ZHE#01

Quant Time: Dec 10 08:12:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

