

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069947.D
 Acq On : 10 Dec 2021 7:10
 Operator : JC/MD
 Sample : M4966-28
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P20-8

Quant Time: Dec 10 08:17:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1108789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2178156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2511160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1035725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	908501	53.068	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.140%	
35) Dibromofluoromethane	8.021	113	669615	49.079	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.160%	
50) Toluene-d8	10.440	98	2893517	52.818	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.640%	
62) 4-Bromofluorobenzene	12.731	95	1212648	61.101	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 122.200%	
Target Compounds						
						Qvalue
16) Acetone	4.291	43	96785	9.398	ug/l	96
18) Methyl Acetate	4.870	43	48621	1.564	ug/l	90
20) Methylene Chloride	5.098	84	139553	10.194	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069947.D
Acq On : 10 Dec 2021 7:10
Operator : JC/MD
Sample : M4966-28
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P20-8

Quant Time: Dec 10 08:17:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

