

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069951.D
 Acq On : 10 Dec 2021 8:52
 Operator : JC/MD
 Sample : M4967-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11S

Quant Time: Dec 10 11:27:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1117905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2194386	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2519960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1034792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	912950	52.893	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.780%	
35) Dibromofluoromethane	8.024	113	670051	48.748	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.500%	
50) Toluene-d8	10.440	98	2909546	52.717	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.440%	
62) 4-Bromofluorobenzene	12.731	95	1217286	60.881	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 121.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	99072	9.542	ug/l	97
18) Methyl Acetate	4.864	43	65022	2.075	ug/l	92
20) Methylene Chloride	5.095	84	135645	9.827	ug/l #	83
43) Isopropyl Acetate	8.566	43	187577	4.077	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069951.D
Acq On : 10 Dec 2021 8:52
Operator : JC/MD
Sample : M4967-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11S

Quant Time: Dec 10 11:27:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

