

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120921\  
 Data File : VN069952.D  
 Acq On : 10 Dec 2021 9:17  
 Operator : JC/MD  
 Sample : M4967-14  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TP-11D

Quant Time: Dec 10 11:27:53 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Dec 03 08:46:47 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.086	168	1131825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2210177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2523648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1046846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	927962	53.101	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.200%
35) Dibromofluoromethane	8.021	113	683599	49.378	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.760%
50) Toluene-d8	10.443	98	2935986	52.816	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.640%
62) 4-Bromofluorobenzene	12.731	95	1220945	60.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.260%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	164330	15.632	ug/l	99
20) Methylene Chloride	5.098	84	128855	9.221	ug/l #	81
25) 2-Butanone	7.340	43	20406	1.429	ug/l	90
43) Isopropyl Acetate	8.568	43	152584	3.293	ug/l #	79
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120921\  
 Data File : VN069952.D  
 Acq On : 10 Dec 2021 9:17  
 Operator : JC/MD  
 Sample : M4967-14  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TP-11D

Quant Time: Dec 10 11:27:53 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Dec 03 08:46:47 2021  
 Response via : Initial Calibration

