

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069953.D
 Acq On : 10 Dec 2021 9:43
 Operator : JC/MD
 Sample : M4965-27
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WHITE-BROWN-CONCRETE-COMP

Quant Time: Dec 10 11:28:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1099011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2143034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2437291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1009787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	903139	53.224	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.440%
35) Dibromofluoromethane	8.024	113	661632	49.289	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.580%
50) Toluene-d8	10.440	98	2821870	52.354	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.700%
62) 4-Bromofluorobenzene	12.731	95	1157068	59.256	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.520%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	91907	9.004	ug/l	93
18) Methyl Acetate	4.862	43	64489	2.093	ug/l #	85
20) Methylene Chloride	5.101	84	43250	3.187	ug/l	90
43) Isopropyl Acetate	8.566	43	182941	4.071	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069953.D
Acq On : 10 Dec 2021 9:43
Operator : JC/MD
Sample : M4965-27
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WHITE-BROWN-CONCRETE-COMP

Quant Time: Dec 10 11:28:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

