

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069904.D
 Acq On : 9 Dec 2021 11:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 10 06:11:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.583	0.565	3.1	83	0.00
3 P	Chloromethane	0.731	0.660	9.7	84	0.00
4 C	Vinyl Chloride	0.672	0.629	6.4#	85	0.00
5 T	Bromomethane	0.365	0.376	-3.0	93	0.00
6 T	Chloroethane	0.412	0.383	7.0	85	0.00
7 T	Trichlorofluoromethane	0.882	0.842	4.5	87	0.00
8 T	Diethyl Ether	0.340	0.339	0.3	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.514	-4.0	95	0.00
10 T	Methyl Iodide	0.675	0.665	1.5	85	0.00
11 T	Tert butyl alcohol	0.146	0.115	21.2	71	-0.01
12 CM	1,1-Dichloroethene	0.476	0.463	2.7#	89	0.00
13 T	Acrolein	0.124	0.151	-21.8	118	0.00
14 T	Allyl chloride	1.073	1.094	-2.0	93	0.00
15 T	Acrylonitrile	0.393	0.366	6.9	83	0.00
16 T	Acetone	0.464	0.467	-0.6	90	0.00
17 T	Carbon Disulfide	1.230	1.244	-1.1	90	0.00
18 T	Methyl Acetate	1.402	1.304	7.0	86	0.00
19 T	Methyl tert-butyl Ether	1.852	1.860	-0.4	91	0.00
20 T	Methylene Chloride	0.617	0.575	6.8	90	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.497	3.9	88	0.00
22 T	Diisopropyl ether	2.098	2.181	-4.0	92	0.00
23 T	Vinyl Acetate	1.707	1.823	-6.8	92	0.00
24 P	1,1-Dichloroethane	1.087	1.103	-1.5	92	0.00
25 T	2-Butanone	0.631	0.587	7.0	82	0.00
26 T	2,2-Dichloropropane	0.845	0.913	-8.0	99	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.612	2.1	88	0.00
28 T	Bromochloromethane	0.565	0.568	-0.5	92	0.00
29 T	Tetrahydrofuran	0.383	0.345	9.9	78	0.00
30 C	Chloroform	1.097	1.099	-0.2#	91	0.00
31 T	Cyclohexane	1.074	1.027	4.4	89	0.00
32 T	1,1,1-Trichloroethane	0.935	0.947	-1.3	90	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.709	8.2	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.313	0.300	4.2	84	0.00
36 T	1,1-Dichloropropene	0.476	0.489	-2.7	89	0.00
37 T	Ethyl Acetate	0.708	0.662	6.5	81	0.00
38 T	Carbon Tetrachloride	0.467	0.503	-7.7	93	0.00
39 T	Methylcyclohexane	0.566	0.605	-6.9	90	0.00
40 TM	Benzene	1.455	1.472	-1.2	88	0.00
41 T	Methacrylonitrile	0.333	0.345	-3.6	87	0.00
42 TM	1,2-Dichloroethane	0.587	0.602	-2.6	90	0.00
43 T	Isopropyl Acetate	1.048	1.069	-2.0	87	0.00
44 TM	Trichloroethene	0.347	0.349	-0.6	87	0.00
45 C	1,2-Dichloropropane	0.391	0.408	-4.3#	92	0.00
46 T	Dibromomethane	0.256	0.262	-2.3	89	0.00
47 T	Bromodichloromethane	0.485	0.547	-12.8	95	0.00
48 T	Methyl methacrylate	0.489	0.532	-8.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
 Data File : VN069904.D
 Acq On : 9 Dec 2021 11:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 10 06:11:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.007	22.2	68	0.00
50 S	Toluene-d8	1.258	1.167	7.2	81	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.678	0.3	82	0.00
52 CM	Toluene	0.888	0.908	-2.3#	87	0.00
53 T	t-1,3-Dichloropropene	0.510	0.593	-16.3	95	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.627	-14.4	93	0.00
55 T	1,1,2-Trichloroethane	0.358	0.368	-2.8	88	0.00
56 T	Ethyl methacrylate	0.608	0.638	-4.9	87	0.00
57 T	1,3-Dichloropropane	0.631	0.647	-2.5	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.162	-5.2	86	0.00
59 T	2-Hexanone	0.514	0.509	1.0	82	0.00
60 T	Dibromochloromethane	0.340	0.396	-16.5	95	0.00
61 T	1,2-Dibromoethane	0.371	0.381	-2.7	87	0.00
62 S	4-Bromofluorobenzene	0.456	0.432	5.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.339	0.327	3.5	84	0.00
65 PM	Chlorobenzene	1.014	1.031	-1.7	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.389	-9.0	91	0.00
67 C	Ethyl Benzene	1.843	1.934	-4.9#	88	0.00
68 T	m/p-Xylenes	0.671	0.709	-5.7	88	0.00
69 T	o-Xylene	0.676	0.702	-3.8	87	0.00
70 T	Styrene	1.072	1.172	-9.3	89	0.00
71 P	Bromoform	0.264	0.316	-19.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	4.449	4.320	2.9	88	0.00
74 T	N-amyl acetate	1.967	1.947	1.0	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.417	7.9	85	0.00
76 T	1,2,3-Trichloropropane	1.434	1.443	-0.6	91	0.00
77 T	Bromobenzene	1.045	0.988	5.5	87	0.00
78 T	n-propylbenzene	4.950	5.042	-1.9	89	0.00
79 T	2-Chlorotoluene	3.128	3.052	2.4	89	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.575	0.2	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.439	-10.9	92	0.00
82 T	4-Chlorotoluene	2.942	2.962	-0.7	90	0.00
83 T	tert-Butylbenzene	3.171	3.098	2.3	87	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.506	-2.8	89	0.00
85 T	sec-Butylbenzene	4.301	4.385	-2.0	89	0.00
86 T	p-Isopropyltoluene	3.362	3.532	-5.1	90	0.00
87 T	1,3-Dichlorobenzene	1.717	1.693	1.4	88	0.00
88 T	1,4-Dichlorobenzene	1.687	1.657	1.8	89	0.00
89 T	n-Butylbenzene	2.727	2.983	-9.4	93	0.00
90 T	Hexachloroethane	0.590	0.669	-13.4	96	0.00
91 T	1,2-Dichlorobenzene	1.673	1.627	2.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.268	3.6	84	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.792	-4.2	92	0.00
94 T	Hexachlorobutadiene	0.443	0.483	-9.0	95	0.00
95 T	Naphthalene	2.199	2.056	6.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120921\
Data File : VN069904.D
Acq On : 9 Dec 2021 11:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 10 06:11:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.763	-1.9	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6