

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
 Data File : VN075677.D
 Acq On : 09 Dec 2022 19:19
 Operator : JC\MD
 Sample : PB149441ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149441ZHE#07

Quant Time: Dec 12 03:11:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 03:01:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	412171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	684026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	619049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	285011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	173682	39.715	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.420%
35) Dibromofluoromethane	8.171	113	179028	42.898	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.800%
50) Toluene-d8	10.571	98	571295	36.451	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	72.900%#
62) 4-Bromofluorobenzene	12.853	95	248688	46.409	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	14837	10.018	ug/l	93
18) Methyl Acetate	5.042	43	5109	1.074	ug/l	99
20) Methylene Chloride	5.283	84	8768	1.986	ug/l	99
37) Ethyl Acetate	7.571	43	23353	5.249	ug/l	98
43) Isopropyl Acetate	8.694	43	114897	15.468	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
Data File : VN075677.D
Acq On : 09 Dec 2022 19:19
Operator : JC\MD
Sample : PB149441ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149441ZHE#07

Quant Time: Dec 12 03:11:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 03:01:58 2022
Response via : Initial Calibration

