

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
 Data File : VN075686.D
 Acq On : 10 Dec 2022 03:35
 Operator : JC\MD
 Sample : N5989-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15

Quant Time: Dec 12 04:24:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 03:01:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	212387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	355643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	318328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	121376	53.862	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.720%	
35) Dibromofluoromethane	8.171	113	110078	50.731	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.460%	
50) Toluene-d8	10.571	98	409598	50.266	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.540%	
62) 4-Bromofluorobenzene	12.853	95	134218	48.174	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.340%	
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	5.806	73	37288	6.092	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
Data File : VN075686.D
Acq On : 10 Dec 2022 03:35
Operator : JC\MD
Sample : N5989-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15

Quant Time: Dec 12 04:24:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 03:01:58 2022
Response via : Initial Calibration

