

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
 Data File : VN075694.D
 Acq On : 10 Dec 2022 06:46
 Operator : JC\MD
 Sample : N5989-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-29

Quant Time: Dec 12 04:56:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 03:01:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	201031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	340102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	304529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	117413	55.046	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.100%	
35) Dibromofluoromethane	8.177	113	103656	49.954	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.900%	
50) Toluene-d8	10.570	98	398391	51.124	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.240%	
62) 4-Bromofluorobenzene	12.853	95	130597	49.017	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	98.040%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.494	96	9305	3.983	ug/l	91
30) Chloroform	7.977	83	3545	0.996	ug/l #	73
38) Carbon Tetrachloride	8.377	117	3466	1.090	ug/l	99
44) Trichloroethene	9.353	130	4957	2.078	ug/l	91
64) Tetrachloroethene	11.106	164	58821	26.518	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
Data File : VN075694.D
Acq On : 10 Dec 2022 06:46
Operator : JC\MD
Sample : N5989-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-29

Quant Time: Dec 12 04:56:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 03:01:58 2022
Response via : Initial Calibration

