

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
 Data File : VN075699.D
 Acq On : 10 Dec 2022 08:45
 Operator : JC\MD
 Sample : N5989-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-32

Quant Time: Dec 12 04:57:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 03:01:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	194369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	331827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	300185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142416	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	111848	54.234	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.460%
35) Dibromofluoromethane	8.177	113	101729	50.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.500%
50) Toluene-d8	10.571	98	384683	50.596	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.200%
62) 4-Bromofluorobenzene	12.853	95	129675	49.884	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120922\
Data File : VN075699.D
Acq On : 10 Dec 2022 08:45
Operator : JC\MD
Sample : N5989-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-32

Quant Time: Dec 12 04:57:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 03:01:58 2022
Response via : Initial Calibration

