

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
 Data File : VN085143.D
 Acq On : 09 Dec 2024 11:11
 Operator : JC\MD
 Sample : VN1209WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1209WBL01

Quant Time: Dec 10 04:57:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	227682	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	197276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	78404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97776	51.377	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.760%		
35) Dibromofluoromethane	8.165	113	82024	53.422	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.840%		
50) Toluene-d8	10.565	98	267585	48.629	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.260%		
62) 4-Bromofluorobenzene	12.847	95	88149	42.837	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 85.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
Data File : VN085143.D
Acq On : 09 Dec 2024 11:11
Operator : JC\MD
Sample : VN1209WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1209WBL01

Quant Time: Dec 10 04:57:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

