

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
 Data File : VN085147.D
 Acq On : 09 Dec 2024 12:57
 Operator : JC\MD
 Sample : P5136-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: Dec 10 05:00:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	276754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	253013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	108724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111316	50.405	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.820%			
35) Dibromofluoromethane	8.165	113	92326	49.469	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.940%			
50) Toluene-d8	10.565	98	333751	49.899	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.800%			
62) 4-Bromofluorobenzene	12.847	95	119874	47.925	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.860%			
Target Compounds						
16) Acetone	4.430	43	74049	115.130	ug/l	97
20) Methylene Chloride	5.271	84	1979	1.008	ug/l #	93
25) 2-Butanone	7.483	43	7582	7.210	ug/l #	81
43) Isopropyl Acetate	8.688	43	94449	19.785	ug/l #	88
52) Toluene	10.629	92	6143	1.298	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
Data File : VN085147.D
Acq On : 09 Dec 2024 12:57
Operator : JC\MD
Sample : P5136-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: Dec 10 05:00:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

