

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
 Data File : VN085148.D
 Acq On : 09 Dec 2024 13:22
 Operator : JC\MD
 Sample : P5095-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-764

Quant Time: Dec 10 05:00:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	141303	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	9.100	114	252983	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	103595	51.469	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.940%	
35) Dibromofluoromethane	8.165	113	84850	49.735	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.480%	
50) Toluene-d8	10.565	98	306637	50.153	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.300%	
62) 4-Bromofluorobenzene	12.847	95	108306	47.369	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.740%	
Target Compounds						
16) Acetone	4.424	43	42634	72.730	ug/l	95
43) Isopropyl Acetate	8.688	43	122083	27.977	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
Data File : VN085148.D
Acq On : 09 Dec 2024 13:22
Operator : JC\MD
Sample : P5095-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-764

Quant Time: Dec 10 05:00:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

