

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
 Data File : VN085150.D
 Acq On : 09 Dec 2024 14:10
 Operator : JC\MD
 Sample : P5096-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-A

Quant Time: Dec 10 05:01:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107246	52.019	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.040%	
35) Dibromofluoromethane	8.165	113	86697	49.001	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.000%	
50) Toluene-d8	10.565	98	316499	49.915	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.840%	
62) 4-Bromofluorobenzene	12.847	95	112572	47.474	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.940%	
Target Compounds						
16) Acetone	4.430	43	43304	72.120	ug/l	Qvalue 99
43) Isopropyl Acetate	8.688	43	175072	38.686	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
Data File : VN085150.D
Acq On : 09 Dec 2024 14:10
Operator : JC\MD
Sample : P5096-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-A

Quant Time: Dec 10 05:01:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

