

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
 Data File : VN085151.D
 Acq On : 09 Dec 2024 14:35
 Operator : JC\MD
 Sample : P5117-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TAPIAL2-IDW-SOIL-120424-00-T2

Quant Time: Dec 10 05:02:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	221770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	100925	53.449	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.900%	
35) Dibromofluoromethane	8.165	113	81241	48.892	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.780%	
50) Toluene-d8	10.565	98	299992	50.377	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.760%	
62) 4-Bromofluorobenzene	12.847	95	99828	44.827	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.660%	
Target Compounds						
16) Acetone	4.430	43	46246	84.094	ug/l	96
25) 2-Butanone	7.495	43	4586	5.101	ug/l	100
43) Isopropyl Acetate	8.689	43	156958	36.930	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120924\
Data File : VN085151.D
Acq On : 09 Dec 2024 14:35
Operator : JC\MD
Sample : P5117-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TAPIAL2-IDW-SOIL-120424-00-T2

Quant Time: Dec 10 05:02:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

