

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120925\
 Data File : VN088481.D
 Acq On : 09 Dec 2025 16:02
 Operator : JC\MD
 Sample : Q3795-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-200059

Quant Time: Dec 10 05:09:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	255958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	593335	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	594444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.765	152	267346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.553	65	270280	56.055	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.100%
35) Dibromofluoromethane	8.147	113	189479	46.082	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	10.541	98	705874	46.138	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.280%
62) 4-Bromofluorobenzene	12.823	95	272982	52.003	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.000%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	144755	83.012	ug/l	97
18) Methyl Acetate	5.012	43	9983	1.923	ug/l #	81
20) Methylene Chloride	5.265	84	295222	79.130	ug/l	98
43) Isopropyl Acetate	8.671	43	31825	2.496	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120925\
Data File : VN088481.D
Acq On : 09 Dec 2025 16:02
Operator : JC\MD
Sample : Q3795-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-200059

Quant Time: Dec 10 05:09:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

