

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052759.D
 Acq On : 10 Dec 2018 17:50
 Operator : MD\SY
 Sample : PB115448ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#08

Quant Time: Dec 11 01:24:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1234190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1902239	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1657043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	623416	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	634722	47.23	ug/l	0.00
Spiked Amount			Recovery	=	94.46%	
35) Dibromofluoromethane	7.59	113	509932	43.48	ug/l	0.00
Spiked Amount			Recovery	=	86.96%	
50) Toluene-d8	10.09	98	2318974	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.40	95	726409	47.36	ug/l	0.00
Spiked Amount			Recovery	=	94.72%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	9972	2.813	ug/l	95
18) Methyl Acetate	4.34	43	5393	0.530	ug/l #	70
20) Methylene Chloride	4.55	84	20047	1.445	ug/l	95
29) Tetrahydrofuran	7.22	42	1426	0.440	ug/l #	69
31) Cyclohexane	7.66	56	21607	0.573	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	83386	4.642	ug/l #	50
38) Carbon Tetrachloride	7.67	117	108157	6.651	ug/l #	16
43) Isopropyl Acetate	8.17	43	76221	4.108	ug/l #	73
45) 1,2-Dichloropropane	9.27	63	3366	0.236	ug/l #	43
48) Methyl methacrylate	9.18	41	14859	1.536	ug/l #	13
49) 1,4-Dioxane	9.27	88	465	4.632	ug/l #	1
51) 4-Methyl-2-Pentanone	10.02	43	17058	1.639	ug/l #	36
54) cis-1,3-Dichloropropene	9.90	75	123615	6.259	ug/l #	62
56) Ethyl methacrylate	10.54	69	1731	2.799	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	25681	1.300	ug/l #	43
59) 2-Hexanone	10.76	43	345850	49.505	ug/l #	51
71) Bromoform	12.40	173	4436	0.564	ug/l #	1
74) N-amyl acetate	11.88	43	42664	3.453	ug/l #	52
76) 1,2,3-Trichloropropane	12.40	75	342325	35.609	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	19488	0.500	ug/l	96
81) trans-1,4-Dichloro-2-buten	12.40	75	342325	118.917	ug/l #	11
93) 1,2,4-Trichlorobenzene	14.91	180	61	2.835	ug/l #	1
95) Naphthalene	15.13	128	662	3.455	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052759.D
Acq On : 10 Dec 2018 17:50
Operator : MD\SY
Sample : PB115448ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#08

Quant Time: Dec 11 01:24:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :


PB115448ZHE#08

Tot Ion:168 Resp: 1234190

Ion Ratio Lower Upper

168 100

99 50.6 41.7 62.5



Abundance Ion 167.90 (167.60 to 168.60): VN052470.D (-1812) (-)

Ion 98.90 (98.60 to 99.60): VN052470.D (-1812) (-)

7.67

600000

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 2.813 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN052759.D


Acq: 10 Dec 2018 17:50

Tgt Ion: 43 Resp: 9972

Ion Ratio Lower Upper

43 100

58 29.9 26.0 39.0



Abundance Ion 43.00 (42.70 to 43.70): VN052470.D (-601) (-)

Ion 58.00 (57.70 to 58.70): VN052470.D (-601) (-)

3.82

4000


3000


2000

1000

0

Time-->









#18
Methyl Acetate
Concen: 0.530 ug/l
RT: 4.34 min Scan# 797
Delta R.T. 0.01 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50




Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#08


Tot Ion: 43 Resp: 5393
Ion Ratio Lower Upper
43 100
74 8.5 18.6 27.8#



#20
Methylene Chloride
Concen: 1.445 ug/l
RT: 4.55 min Scan# 864
Delta R.T. -0.00 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50

Tgt Ion: 84 Resp: 20047
Ion Ratio Lower Upper
84 100
49 128.3 98.0 147.0
51 40.1 30.1 45.1
86 66.4 50.7 76.1





#29

Tetrahydrofuran

Concen: 0.440 ug/l

RT: 7.22 min Scan# 1693

Delta R.T. 0.01 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08





Tot Ion: 42 Resp: 1426

Ion Ratio Lower Upper

42 100

72 16.4 33.3 49.9#

71 26.2 32.0 48.0#



#31

Cyclohexane

Concen: 0.573 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. 0.01 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50




Tgt Ion: 56 Resp: 21607


Ion Ratio Lower Upper

56 100

69 171.8 25.8 38.6#

84 143.3 65.3 97.9#









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. 0.00 min
 Lab File: VN052759.D
 Acq: 10 Dec 2018 17:50




Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#08


Tot Ion: 65 Resp: 634722
 Ion Ratio Lower Upper
 65 100
 67 54.5 0.0 109.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN052759.D
 Acq: 10 Dec 2018 17:50

Tgt Ion: 114 Resp: 1902239
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 41.8
 88 15.2 0.0 31.6









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN052759.D
 Acq: 10 Dec 2018 17:50




Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#08


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.1	82.0	123.0
192	22.1	17.0	25.6



#36
 1,1-Dichloropropene
 Concen: 4.642 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.13 min
 Lab File: VN052759.D
 Acq: 10 Dec 2018 17:50

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.7	17.5	52.6#
77	0.0	24.5	36.7#





#38

Carbon Tetrachloride

Concen: 6.651 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.11 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion:117 Resp: 108157

Ion Ratio Lower Upper

117 100

119 5.1 76.8 115.2#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 4.108 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50


Tgt Ion: 43 Resp: 76221

Ion Ratio Lower Upper

43 100

61 12.3 24.1 36.1#



87 7.5 10.6 15.8#




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V





#45

1,2-Dichloropropane

Concen: 0.236 ug/l

RT: 9.27 min Scan# 2331

Delta R.T. 0.15 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :


PB115448ZHE#08

Tot Ion: 63 Resp: 3366

Ion Ratio Lower Upper




63 100

65 0.0 25.0 37.6#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0



#48

Methyl methacrylate

Concen: 1.536 ug/l

RT: 9.18 min Scan# 2303

Delta R.T. -0.02 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50


Tgt Ion: 41 Resp: 14859

Ion Ratio Lower Upper

41 100

69 0.0 70.0 105.0#



39 0.0 42.6 64.0#




Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 39.00 (38.70 to 39.70): VN0





#49

1,4-Dioxane

Concen: 4.632 ug/l

RT: 9.27 min Scan# 2331

Delta R.T. 0.07 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion: 88 Resp: 465

Ion Ratio Lower Upper

88 100

43 563403.2 26.4 39.6#




58 2543.7 55.4 83.0#



Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 4.632 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052759.D


Acq: 10 Dec 2018 17:50

Tgt Ion: 98 Resp: 2318974

Ion Ratio Lower Upper



98 100


100 64.0 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.639 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :





PB115448ZHE#08

Tot Ion: 43 Resp: 17058

Ion Ratio Lower Upper

43 100

58 0.0 31.5 47.3#



#54

cis-1,3-Dichloropropene

Concen: 6.259 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.06 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50




Tgt Ion: 75 Resp: 123615


Ion Ratio Lower Upper

75 100

77 0.5 26.2 39.2#

39 29.4 36.0 54.0#





#56

Ethyl methacrylate

Concen: 2.799 ug/l

RT: 10.54 min Scan# 2728

Delta R.T. 0.12 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion: 69 Resp: 1731

Ion Ratio Lower Upper

69 100

41 5541.7 52.6 79.0#




39 2177.1 25.9 38.9#



Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 1.300 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.05 min

Lab File: VN052759.D


Acq: 10 Dec 2018 17:50

Tgt Ion: 76 Resp: 25681

Ion Ratio Lower Upper



76 100


78 0.3 25.6 38.4#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 49.505 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50




Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#08


Tot Ion: 43 Resp: 345850
Ion Ratio Lower Upper
43 100
58 19.2 27.4 82.0#



#62
4-Bromofluorobenzene
Concen: N.D. ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50

Tgt Ion: 95 Resp: 726409
Ion Ratio Lower Upper
95 100
174 88.4 0.0 173.0
176 84.9 0.0 166.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#08

Tot Ion:117 Resp: 1657043
Ion Ratio Lower Upper
117 100
82 53.9 43.6 65.4
119 32.3 25.4 38.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.564 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.28 min
Lab File: VN052759.D
Acq: 10 Dec 2018 17:50

Tgt Ion:173 Resp: 4436
Ion Ratio Lower Upper
173 100
175 906.7 24.4 73.2#
254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion:152 Resp: 623416

Ion Ratio Lower Upper

152 100

115 56.4 28.4 85.2

150 157.7 0.0 346.0






Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 3.453 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Tgt Ion: 43 Resp: 42664


Ion Ratio Lower Upper

43 100

70 1.5 33.8 50.8#

55 4.6 20.0 30.0#

61 10.1 19.7 29.5#





Abundance


Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 35.609 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :


PB115448ZHE#08

Tot Ion: 75 Resp: 342325

Ion Ratio Lower Upper

75 100

77 1.4 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN052759.D (-3305) (-)

Ion 77.00 (76.70 to 77.70): VN052759.D (-3305) (-)

250000

200000

150000



100000

50000

0

12.35 12.40 12.45

Time-->



#80

1,3,5-Trimethylbenzene

Concen: 0.500 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN052759.D


Acq: 10 Dec 2018 17:50

Tgt Ion: 105 Resp: 19488

Ion Ratio Lower Upper

105 100

120 51.6 24.3 73.0



Abundance Ion 105.00 (104.70 to 105.70): VN052759.D (-3316) (-)

Ion 120.00 (119.70 to 120.70): VN052759.D (-3316) (-)


10000


5000

0

12.70 12.75

Time-->





#81

trans-1,4-Dichloro-2-butene

Concen: 118.917 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion: 75 Resp: 342325

Ion Ratio Lower Upper

75 100

53 0.0 77.4 116.0#




89 0.0 35.4 53.2#



Abundance Ion 74.90 (74.60 to 75.60): VN052759.D (-3181) (-)

Ion 53.00 (52.70 to 53.70): VN052759.D (-3181) (-)

Ion 88.90 (88.60 to 89.60): VN052759.D (-3181) (-)



#93

1,2,4-Trichlorobenzene

Concen: 2.835 ug/l

RT: 14.91 min Scan# 4086

Delta R.T. 0.00 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50


Tgt Ion: 180 Resp: 61

Ion Ratio Lower Upper

180 100

182 231.1 47.8 143.4#



145 0.0 14.1 42.3#




Abundance Ion 179.80 (179.50 to 180.50): VN052759.D (-3993) (-)

Ion 181.80 (181.50 to 182.50): VN052759.D (-3993) (-)

Ion 144.80 (144.50 to 145.50): VN052759.D (-3993) (-)





#95

Naphthalene

Concen: 3.455 ug/l

RT: 15.13 min Scan# 4154

Delta R.T. -0.01 min

Lab File: VN052759.D

Acq: 10 Dec 2018 17:50

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#08


Tot Ion:128 Resp: 662

Ion Ratio Lower Upper

128 100

127 0.0 10.2 15.2#

129 0.0 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

