

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052752.D
 Acq On : 10 Dec 2018 14:43
 Operator : MD\SY
 Sample : PB115448ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#01

Quant Time: Dec 11 02:34:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1211137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1880883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1656502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	663600	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	635112	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
35) Dibromofluoromethane	7.59	113	497041	42.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.72%	
50) Toluene-d8	10.09	98	2297194	50.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.78%	
62) 4-Bromofluorobenzene	12.40	95	735767	48.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.02%	
Target Compounds						
16) Acetone	3.82	43	15807	4.545	ug/l	96
20) Methylene Chloride	4.56	84	67999	4.996	ug/l	97
43) Isopropyl Acetate	8.17	43	65094	3.548	ug/l #	76
95) Naphthalene	15.13	128	3938	3.575	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052752.D
Acq On : 10 Dec 2018 14:43
Operator : MD\SY
Sample : PB115448ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#01

Quant Time: Dec 11 02:34:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

