

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052753.D
 Acq On : 10 Dec 2018 15:10
 Operator : MD\SY
 Sample : PB115448ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#02

Quant Time: Dec 11 02:37:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1239051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1899503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1663653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	665507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	639415	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
35) Dibromofluoromethane	7.59	113	503369	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
50) Toluene-d8	10.09	98	2333983	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	12.40	95	744433	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
Target Compounds						
16) Acetone	3.82	43	16625	4.672	ug/l	91
20) Methylene Chloride	4.55	84	68619	4.928	ug/l	98
43) Isopropyl Acetate	8.17	43	78212	4.222	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052753.D
Acq On : 10 Dec 2018 15:10
Operator : MD\SY
Sample : PB115448ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#02

Quant Time: Dec 11 02:37:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

