

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052755.D
 Acq On : 10 Dec 2018 16:03
 Operator : MD\SY
 Sample : PB115448ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#04

Quant Time: Dec 11 02:45:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1224899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1891690	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1668729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	657921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	638136	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
35) Dibromofluoromethane	7.59	113	500957	42.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.92%	
50) Toluene-d8	10.09	98	2322021	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.40	95	732623	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
Target Compounds						
16) Acetone	3.82	43	15794	4.490	ug/l	96
20) Methylene Chloride	4.56	84	68351	4.965	ug/l	97
43) Isopropyl Acetate	8.17	43	70694	3.832	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052755.D
Acq On : 10 Dec 2018 16:03
Operator : MD\SY
Sample : PB115448ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#04

Quant Time: Dec 11 02:45:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

