

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052758.D
 Acq On : 10 Dec 2018 17:23
 Operator : MD\SY
 Sample : PB115448ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#07

Quant Time: Dec 11 02:52:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1185026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1840605	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1641766	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	689161	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	614880	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
35) Dibromofluoromethane	7.59	113	490221	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	
50) Toluene-d8	10.09	98	2246275	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.40	95	740265	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
Target Compounds						
16) Acetone	3.82	43	11363	3.339	ug/l	98
20) Methylene Chloride	4.55	84	21436	1.610	ug/l	93
43) Isopropyl Acetate	8.17	43	68360	3.808	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052758.D
Acq On : 10 Dec 2018 17:23
Operator : MD\SY
Sample : PB115448ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#07

Quant Time: Dec 11 02:52:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

