

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052761.D
 Acq On : 10 Dec 2018 18:43
 Operator : MD\SY
 Sample : PB115448ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#10

Quant Time: Dec 11 02:58:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1186247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1839923	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1648476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	700489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	629495	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
35) Dibromofluoromethane	7.59	113	489642	43.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.34%	
50) Toluene-d8	10.09	98	2258689	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.40	95	747671	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
Target Compounds						
16) Acetone	3.82	43	12468	3.660	ug/l	91
20) Methylene Chloride	4.56	84	20824	1.562	ug/l	96
43) Isopropyl Acetate	8.17	43	77172	4.301	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052761.D
Acq On : 10 Dec 2018 18:43
Operator : MD\SY
Sample : PB115448ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#10

Quant Time: Dec 11 02:58:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

