

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052775.D
 Acq On : 11 Dec 2018 2:17
 Operator : MD\SY
 Sample : J6196-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-03-120718-B

Quant Time: Dec 11 03:37:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1274475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1968225	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1780521	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	763276	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	662501	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
35) Dibromofluoromethane	7.59	113	560296	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
50) Toluene-d8	10.09	98	2403464	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
62) 4-Bromofluorobenzene	12.40	95	811859	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
Target Compounds						
16) Acetone	3.82	43	31589	8.631	ug/l	98
20) Methylene Chloride	4.55	84	6202800	433.064	ug/l	99
43) Isopropyl Acetate	8.17	43	63664	3.317	ug/l #	76
95) Naphthalene	15.13	128	34774	4.556	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052775.D
Acq On : 11 Dec 2018 2:17
Operator : MD\SY
Sample : J6196-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EO-03-120718-B

Quant Time: Dec 11 03:37:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

