

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052776.D
 Acq On : 11 Dec 2018 2:44
 Operator : MD\SY
 Sample : J6196-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-03-120718-C

Quant Time: Dec 11 03:39:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1214323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1876249	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1636707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	650660	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	637384	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	7.59	113	532535	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
50) Toluene-d8	10.09	98	2308517	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.40	95	726238	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
Target Compounds						
16) Acetone	3.82	43	97842	28.056	ug/l	98
20) Methylene Chloride	4.55	84	2871098	210.382	ug/l	98
43) Isopropyl Acetate	8.17	43	65490	3.579	ug/l #	74
84) 1,2,4-Trimethylbenzene	13.04	105	48370	1.185	ug/l	100
95) Naphthalene	15.13	128	209442	11.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052776.D
Acq On : 11 Dec 2018 2:44
Operator : MD\SY
Sample : J6196-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EO-03-120718-C

Quant Time: Dec 11 03:39:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

