

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052784.D
 Acq On : 11 Dec 2018 6:18
 Operator : MD\SY
 Sample : J6316-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BORING-SAMPLE-CBH-591

Quant Time: Dec 11 07:33:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1192698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1849690	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1614096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	634279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	626008	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	7.59	113	535790	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	10.09	98	2234732	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.40	95	704595	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
Target Compounds						
16) Acetone	3.82	43	18345	5.356	ug/l	97
20) Methylene Chloride	4.56	84	33216	2.478	ug/l	96
43) Isopropyl Acetate	8.17	43	79087	4.384	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052784.D
Acq On : 11 Dec 2018 6:18
Operator : MD\SY
Sample : J6316-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BORING-SAMPLE-CBH-591

Quant Time: Dec 11 07:33:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

